

Certificate of Analysis

Cat.no. 5651808
Product CASYton
Lot 3020926

22.09.2026
Page 01 of 01

Parameter	Unit	Target Value	Result
pH-value		7.3 – 7.5	7.4
Osmolality	mOs/kg H2O	295 - 325	311
Conductivity	mS/cm	15.4 – 16.0	15.7
Particle purity			corresponds
Performance			corresponds

This lot meets our specification.

Expiry date: 09/2028

OMNI Life Science GmbH & Co. KG



(Dr. Josef Freund)
Head of Production